

Minutes of the Pre-bid meeting held on 4-9-2013 at TEQIP Office, Coimbatore Institute of Technology, Coimbatore – 641 014.

The following member from CIT were present:


1. Dr. N. Murugan, Professor of Mechanical Engineering.
2. Dr. V. Manikandan, Associate Professor of Electrical engineering

The representatives from the following companies were present:

1. Mr. S. Saravanan, Senior Manager, Carl Zeiss
2. Mr. P. R. Mohan, General Manager, Jeol
3. Mr. Vijai Chandrasekaran, Marketing Executive, Advanced Scientific Equipment Pvt. Ltd.

The details provided for the clarifications sought by the members are given below.

1. For the wave length dispersive spectrometer, the crystals to be covered are Lithium fluoride, Pet, Tap and LSM080.
2. The variable temperature (-50 to 1000 °C) is optional
3. Warranty will include two additional filaments along with the consumables to be supplied.
4. The variable vacuum detector is optional.
5. Column size may be increased to accommodate more no. of ports to facilitate upgradation.
6. Minimum eligibility of the bidder required to participate in bidding is as per NPIU norms provided in the bid document.


TEQIP Coordinator 6/9/13

Coordinator
TEQIP Phase II-CIT


6/9/2013
Dr. V. SENTHADURAI, M.E., Ph.D
FIE., FII.E., Sr.M (ORSI) LM (SSI)
PRINCIPAL
COIMBATORE INSTITUTE OF TECHNOLOGY
COIMBATORE - 641 014 INDIA